

OI-922

Page 29 of 67

Rev. BN

Vialed and Packaged Test Strip Release

Test Strips

*nova*  
biomedical

**CERTIFICATE OF COMPLIANCE**

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	N/A
Description:	STATSENSOR-I CREAT Test Strips
Catalog Number:	P/N 43272
Nova Biomedical P/N	P/N 43272
Control Level 1: CV (%) $\leq$ 15% Control Level 2: CV (%) $\leq$ 10% Control Level 3: CV (%) $\leq$ 10%	PASS / FAIL PASS / FAIL PASS / FAIL
Qty:	137
Lot No.	4925058129
Mfg. Date:	2025-02-27
Expiration Date:	2026-02-28
Unique Device Identifier (UDI):	(01)20385480432720 (11) 250227 (17) 260228 (10) 4925058129

Mark Brown  
Quality Control Manager

4/4/25  
Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454