



CERTIFICATE OF ANALYSIS

STP Prime Plus Vet Reference Cartridge
Lot Number 25100062

Ref Number 59345
Exp Date 2026-10-10
Date of Manufacture 2025-04-10

Internal QC Level 4	<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	138.0 - 146.0	Pass
K mmol/L	3.71 - 4.21	Pass
CL mmol/L	122.8 - 131.8	Pass

External QC Level 4	<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	134.3 - 142.3	Pass
K mmol/L	3.73 - 4.23	Pass
CL mmol/L	118.9 - 127.9	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi M

Approval

QC Inspector

Title

4/29/2025

Date