

StatPro pHOx/CCX SO2 Calibrator

Ref Number 22771

Lot Number 25085054

Exp Date 2026-09-18 Date of Manufacture 2025-03-18 UDI# (01)00385480227711(11)250318(17)260918(10)25085054

Level 1 L/N: 25077059 Level 1 Exp Date: 2026-09-18

Pass/Fail <u>Spec</u> 45 - 65 SO2 % Pass Pass Hct % 51 - 55

Level 2 L/N: 25077060 Level 2 Exp Date: 2026-09-18 Pass/Fail <u>Spec</u>

SO2 % 92 - 105 Pass Hct % 51 - 55 Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials. SO2 is traced to tonometry.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi. M Approval

QC Inspector Title

4/30/2025 Date