



## CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 400 Sample  
Lot Number 25112060

Ref Number 53463  
Exp Date 2026-09-28  
Date of Manufacture 2025-03-28  
UDI# (01)00385480534635(11)250328(17)260928(10)25112060

Cal A	L/N	25091016	
		<u>Assigned Value*</u>	<u>Result</u>
	pH	7.356	N/A
pCO2	mmHg	25.0	N/A

Cal B	L/N	25094019	
		<u>Spec</u>	<u>Pass/Fail</u>
	pH	6.837 - 6.843	Pass
pO2	mmHg	167.0	N/A

Flush	L/N	25087023	
		<u>Assigned Value*</u>	<u>Result</u>
pCO2	mmHg	51.0	N/A
pO2	mmHg	98.8	N/A

Reference	L/N	25094020	
		<u>Spec</u>	<u>Pass/Fail</u>
KCl		1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Auishi M*  
Approval

QC Inspector  
Title

04/30/2025  
Date