



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 100 Sample
Lot Number 25112060

Ref Number 53104
Exp Date 2026-09-28
Date of Manufacture 2025-03-28
UDI# (01)00385480531047(11)250328(17)260928(10)25112060

Cal A L/N 25091016			
	<u>Assigned Value*</u>	<u>Result</u>	
pH	7.356	N/A	
pCO2 mmHg	25.0	N/A	

Cal B L/N 25094019			
	<u>Spec</u>	<u>Pass/Fail</u>	
pH	6.837 - 6.843	Pass	
pO2 mmHg	167.0	N/A	

Flush L/N 25087023			
	<u>Assigned Value*</u>	<u>Result</u>	
pCO2 mmHg	51.0	N/A	
pO2 mmHg	98.8	N/A	

Reference L/N 25094020			
	<u>Spec</u>	<u>Pass/Fail</u>	
KCl	1.8 - 2.2	Pass	

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector
Title

05/05/2025
Date