



CERTIFICATE OF ANALYSIS

STP Prime CCS Comp Cal Cartridge 500 Sample  
Lot Number 25105033

Ref Number 53469  
Exp Date 2026-09-18  
Date of Manufacture 2025-03-18  
UDI# (01)00385480534697(11)250318(17)260918(10)25105033

Cal A	L/N	25084073	
		<u>Assigned Value*</u>	<u>Result</u>
	pH	7.354	N/A
	pCO2 mmHg	25.0	N/A
		<u>Spec</u>	<u>Pass/Fail</u>
	Na mmol/L	132.1 - 133.9	Pass
	K mmol/L	3.97 - 4.03	Pass
	Cl mmol/L	105.2 - 106.8	Pass
	iCa mmol/L	1.08 - 1.12	Pass
	Glu mg/dL	79 - 81	Pass
	Lac mmol/L	1.8 - 2.2	Pass

Cal B	L/N	25084074	
		<u>Spec</u>	<u>Pass/Fail</u>
	pH	6.837 - 6.843	Pass
	pO2 mmHg	167.0	N/A
	Na mmol/L	68.0 - 69.0	Pass
	K mmol/L	9.92 - 10.08	Pass
	Cl mmol/L	45.6 - 46.4	Pass
	iCa mmol/L	2.17 - 2.23	Pass
	Glu mg/dL	197 - 203	Pass
	Lac mmol/L	9.5 - 10.5	Pass

Flush	L/N	25077046	
		<u>Assigned Value*</u>	<u>Result</u>
	pCO2 mmHg	50.7	N/A
	pO2 mmHg	99.8	N/A

Reference	L/N	25085016	
		<u>Spec</u>	<u>Pass/Fail</u>
	KCl	1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector  
Title

05/06/2025  
Date