



## CERTIFICATE OF ANALYSIS

STP pHox Plus Chloride Sensor  
Lot Number 25111041

Part Number 27812  
Exp Date 2026-07-21  
Date of Manufacture 2025-04-21  
UDI# (01)10385480278123(11)250421(17)260721(10)25111041

	<u>Spec*</u>	<u>Pass/Fail</u>
Slope	$\geq 9.00$	Pass
Std A mV	55.00 - 95.00	Pass
pHox Ctl L1	120.0 - 129.0	Pass
pHox Ctl L3	79.6 - 89.6	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "C. S. S. S.", written over a horizontal line.

Approval

QC Inspector  
Title

5/5/2025  
Date