



CERTIFICATE OF ANALYSIS

STP Prime Plus Reference Cartridge
Lot Number 25101102

Ref Number 57823

Exp Date 2026-10-11

Date of Manufacture 2025-04-11

UDI# (01)00385480578233(11)250411(17)261011(10)25101102

Internal QC Level 4	<u>Spec</u>	<u>Pass/Fail</u>
Na, mmol/L	138.0 - 146.0	Pass
K, mmol/L	3.71 - 4.21	Pass
CL, mmol/L	122.8 - 131.8	Pass
External QC Level 4	<u>Spec</u>	<u>Pass/Fail</u>
Na, mmol/L	134.3 - 142.3	Pass
K, mmol/L	3.73 - 4.23	Pass
CL, mmol/L	118.9 - 127.9	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2016, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

5/6/2025

Date