



## CERTIFICATE OF ANALYSIS

STP® pHox Ultra™ / CCX™ Ionized Mg Sensor  
Lot Number 25104033

Part Number 48556  
Exp Date 2026-07-14

Date of Manufacture 2025-04-14

UDI# (01)10385480485569(11)250414(17)260714(10)25104033

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.70 - 14.00	Pass
Std C mV	-100.00 - +60.00	Pass
pHox Ctl L4	0.55 - 0.73	Pass
pHox Ctl L5	1.00 - 1.30	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

4/30/2025  
Date