



**CERTIFICATE OF ANALYSIS**

Nipro Glucose and  $\beta$ -Ketone Control Solution Level 3  
Lot Number 0425097303

Ref Number 11-117  
Exp Date 2027.10.07

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	275 - 325	300
Ketone mmol/L	2.5 - 3.7	3.2

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

5/8/2025

Date