



CERTIFICATE OF ANALYSIS

STAT PROFILE® pHox Ultra™/CCX™ pCO2 Sensor
Lot Number 25105049

Ref Number 48600
Exp Date 2027-04-15

Date of Manufacture 2025-04-15

UDI# (01)00385480486002(11)250415(17)270415(10)25105049

| | <u>Spec</u> | <u>Pass/Fail</u> |
|----------|---------------|------------------|
| Slope | 9.40 - 11.60 | Pass |
| Std C mV | 0.00 ± 150.00 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to be "John", written in a cursive style.

Approval

QC Inspector

Title

5/7/2025

Date