



CERTIFICATE OF ANALYSIS

Bioprofile Control Level 5
Lot Number 508368

Catalog Number 24619
Exp Date 2026-09

| | <u>Spec</u> | <u>Pass/Fail</u> |
|------------|--------------|------------------|
| Gln mmol/L | 4.60 - 5.40 | Pass |
| Glu mmol/L | 1.80 - 2.20 | Pass |
| Gluc g/L | 9.60 - 10.60 | Pass |
| Lac g/L | 4.60 - 5.40 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP) and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Lac, Gluc.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to read "M. Houghton", with a long, sweeping horizontal line extending to the right.

Approval

QC Inspector
Title

5/8/2025
Date