



CERTIFICATE OF ANALYSIS

STP® pHox Ultra™ / CCX™ Lactate Sensor
Lot Number 25127022

Part Number 48553
Exp Date 2027-05-07

Date of Manufacture 2025-05-07

UDI# (01)00385480485531(11)250507(17)270507(10)25127022

| | <u>Spec</u> | <u>Pass/Fail</u> |
|-------------|---------------|------------------|
| Slope | 11.00 - 45.00 | Pass |
| pHox Ctl L3 | 75 - 89 | Pass |
| Set A L1 | ≥ 470 | Pass |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

5/15/2025
Date