



CERTIFICATE OF ANALYSIS

Allegro Control 2 UACR
Lot Number 8525083242

Exp Date 2027-03-24
Date of Manufacture 2025-03-24
UDI# (01)00385480510318(11)250324(17)270324(10)8525083242

Ref Number 51031

	Spec	Pass/Fail
Albumin, mg/dL	127-193	Pass
Creatinine, mg/dL	196-288	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

5/14/2025
Date