



CERTIFICATE OF ANALYSIS

STP Prime Plus Calibrator Cartridge 100 Sample
Lot Number 25119065

Ref Number 57825
Exp Date 2026-10-04

Date of Manufacture 2025-04-04

UDI# (01)00385480578257(11)250404(17)261004(10)25119065

Table with 3 columns: Cal A L/N 25113009, Cal B L/N 25113013, Cal C L/N 25113014. Rows include analytes like HCO3, Na, K, Cl, Ca, Mg, Glu, Lac, BUN with Spec and Pass/Fail columns. Includes a row for tHb g/dL with Assigned Value* and Result.

Table with 3 columns: Cal D L/N 25113015, Flush L/N 25094011, Reference L/N 25113016. Rows include pH @ 37C, PO2, KCl with Spec and Pass/Fail columns. Includes a row for Assigned Value* and Result.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector
Title

5/16/2025
Date