



CERTIFICATE OF ANALYSIS

STP® pHox Ultra™ / CCX™ BUN Sensor
Lot Number 25126049

Part Number 48555
Exp Date 2026-08-06
Date of Manufacture 2025-05-06
UDI# (01)00385480485555(11)250506(17)260806(10)25126049

| | <u>Spec</u> | <u>Pass/Fail</u> |
|-------------|------------------|------------------|
| Slope | ≥ 9.60 | Pass |
| pHox Ctl L4 | 14 - 24 | Pass |
| pHox Ctl L5 | 49 - 63 | Pass |
| Std D mV | 0.00 ± 60.00 | Pass |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

5/19/2025

Date