



CERTIFICATE OF ANALYSIS

Prime Cal Cartridge ES Comp 500 Sample
Lot Number 25125079

Ref Number 53461
Exp Date 2026-10-14
Date of Manufacture 2025-04-14
UDI# (01)00385480534611(11)250414(17)261014(10)25125079

| Cal A L/N 25105035 | | | Cal B L/N 25105036 | | | Reference L/N 25104036 | | |
|--------------------|---------------|------------------|--------------------|--------------|------------------|------------------------|-------------|------------------|
| | <u>Spec</u> | <u>Pass/Fail</u> | | <u>Spec</u> | <u>Pass/Fail</u> | | <u>Spec</u> | <u>Pass/Fail</u> |
| Na mmol/L | 134.1 - 135.9 | Pass | Na mmol/L | 76.5 - 77.5 | Pass | KCl mol/L | 1.8 - 2.2 | Pass |
| K mmol/L | 4.02 - 4.08 | Pass | K mmol/L | 9.80 - 10.20 | Pass | | | |
| Cl mmol/L | 114.1 - 115.9 | Pass | Cl mmol/L | 44.6 - 45.5 | Pass | | | |
| iCa mmol/L | 0.99 - 1.01 | Pass | iCa mmol/L | 1.98 - 2.02 | Pass | | | |
| iMg mmol/L | 0.36 - 0.46 | Pass | iMg mmol/L | 1.48 - 1.52 | Pass | | | |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

05/16/2025
Date