



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 300 Sample  
Lot Number 25129065

Ref Number 53359  
Exp Date 2026-10-25  
Date of Manufacture 2025-04-25  
UDI# (01)00385480533591(11)250425(17)261025(10)25129065

Cal A	L/N	25115068	
		<u>Assigned Value*</u>	<u>Result</u>
	pH	7.353	N/A
	pCO2 mmHg	25.2	N/A

Cal B	L/N	25115069	
		<u>Spec</u>	<u>Pass/Fail</u>
	pH	6.837 - 6.843	Pass
	pO2 mmHg	167.0	N/A

Flush	L/N	25115045	
		<u>Assigned Value*</u>	<u>Result</u>
	pCO2 mmHg	50.2	N/A
	pO2 mmHg	99.5	N/A

Reference	L/N	25115024	
		<u>Spec</u>	<u>Pass/Fail</u>
	KCl	1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

05/19/2025  
Date