



CERTIFICATE OF ANALYSIS

STP Prime+ Vet Cal-Cartridge 200 Sample w. Creat/BUN
Lot Number 25119066

Ref Number 58405
Exp Date 2026-10-04
Date of Manufacture 2025-04-04

Table with 3 columns: Cal A L/N 25113009, Cal B L/N 25113013, Cal C L/N 25113014. Rows include HCO3, Na, K, Cl, Ca, Mg, Glu, Lac, BUN, Creat and tHb g/dL.

Table with 3 columns: Cal D L/N 25113015, Flush L/N 25094011, Reference L/N 25113016. Rows include pH @ 37C and KCl.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector
Title

5/21/2025
Date