



CERTIFICATE OF ANALYSIS

STP pHox /Basic/Plus/L/C/M pCO2 Sensor
Lot Number 25127024

Ref Number 21524

Exp Date 2027-05-07

Date of Manufacture 2025-05-07

UDI (01)00385480215244(11)250507(17)270507(10)25127024

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.50 - 11.30	Pass
Std A mV	0.00 +/- 150.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

A handwritten signature in dark ink, appearing to be a stylized 'J' followed by a series of loops and a horizontal stroke.

Approval

QC Inspector

Title

5/22/2025

Date