



## CERTIFICATE OF ANALYSIS

Stat Profile® pHox Ultra™ / CCX™ SO2 Sensor  
Lot Number 25135006

Part Number 48591  
Exp Date 2028-05-15  
Date of manufacture 2025-05-15  
UDI#(01)00385480485913(11)250515(17)280515(10)25135006

|       | <u>Spec</u>  | <u>Mean Result</u> | <u>Pass/Fail</u> |
|-------|--------------|--------------------|------------------|
| Slope | 7.00 - 14.00 | 9.24               | Pass             |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to be "J. Smith", written over a faint horizontal line.

Approval

QC Inspector  
Title

5/20/2025  
Date