



**CERTIFICATE OF COMPLIANCE**

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA	
Description:	TEST STRIPS, STATSTRIP™ GLUCOSE, PACKAGED 1800 COUNT	
Catalog Number:	42214	
Nova Biomedical P/N	42214	
Target Range (mg/dL)	SD(mg/dL) / CV (%)	
30-50	≤ 4 mg/dL	PASS / FAIL
51-110	≤ 6.0%	PASS / FAIL
111-150	≤ 5.0%	PASS / FAIL
151-250	≤ 5.0%	PASS / FAIL
251-400	≤ 5.0%	PASS / FAIL
Accuracy Criteria 1: Total within +/-12 mg/dL or +/- 12% of reference method ≥ 95%		PASS / FAIL
Accuracy Criteria 2: Total within +/-15 mg/dL or +/- 15% of reference method ≥ 98%		PASS / FAIL
Qty:	9,450	
Lot No.	0325134249	
Mfg. Date:	2025-05-14	
Expiration Date:	2027-05-14	
Unique Device Identifier (UDI):	(01)20385480422141 (11) 250514 (17) 270514 (10) 0325134249	

*Marc Browne*  
 Quality Control Manager

5/28/25  
 Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454