



CERTIFICATE OF ANALYSIS

STP Prime Plus VET Cal-Cartridge 500 Sample
Lot Number 25141043

Ref Number 58396
Exp Date 2026-11-05
Date of Manufacture 2025-05-05

Table with 3 columns: Cal A L/N 25125070, Cal B L/N 25125071, Cal C L/N 25125072. Rows include HCO3 mmol/L, Na mmol/L, K mmol/L, Cl mmol/L, Ca mmol/L, Mg mmol/L, Glu mg/dL, Lac mmol/L, BUN mg/dL, pH, tHb g/dL, and Assigned Value*/Result.

Table with 3 columns: Cal D L/N 25125073, Flush L/N 25125034, Reference L/N 25125069. Rows include pH @ 37C, PO2, and KCl with Spec and Pass/Fail columns.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

5/28/2025
Date