



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 400 Sample  
Lot Number 25142039

Ref Number 53463  
Exp Date 2026-11-02  
Date of Manufacture 2025-05-02  
UDI# (01)00385480534635(11)250502(17)261102(10)25142039

Cal A	L/N	25122035	
		<u>Assigned Value*</u>	<u>Result</u>
pH		7.355	N/A
pCO2 mmHg		25.1	N/A

Cal B	L/N	25122036	
		<u>Spec</u>	<u>Pass/Fail</u>
pH		6.837 - 6.843	Pass
pO2 mmHg		167.0	N/A

Flush	L/N	25122024	
		<u>Assigned Value*</u>	<u>Result</u>
pCO2 mmHg		51.5	N/A
pO2 mmHg		100.9	N/A

Reference	L/N	25122029	
		<u>Spec</u>	<u>Pass/Fail</u>
KCl		1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

  
Approval

QC Inspector  
Title

05/30/2025  
Date