



CERTIFICATE OF ANALYSIS

STP® pHOx Ultra™ / CCX™ Creatinine Sensor
Lot Number 25141005

Part Number 48595
Exp Date 2027-05-21
Date of Manufacture 2025-05-21
UDI# (01)00385480485951(11)250521(17)270521(10)25141005

| | <u>Specs</u> | <u>Pass/Fail</u> |
|-------------|--------------|------------------|
| Slope | ≥ 5.00 | Pass |
| pHOx Ctl L4 | 0.7 - 1.3 | Pass |
| pHOx Ctl L5 | 5.2 - 7.2 | Pass |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

06/02/2025

Date