



**CERTIFICATE OF ANALYSIS**

Nova CRT 16™ Creatinine Membrane Kit  
Lot Number 25156033

Part Number 28241  
Exp Date 2026-09-05  
Date of manufacture 2025-06-05  
UDI# (01)00385480282413(11)250605(17)260905(10)25156033

	<u>Spec*</u>	<u>Mean Result</u>
Slope	> 45.0	126.6
Std A mV	< 500.00	350.87

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

06/10/2025  
Date