nova biomedical

CERTIFICATE OF ANALYSIS

Allegro Lipids Test Cartridge
Lot Number 25153017

Ref Number 54652
Exp Date 2026-12-02
Date of Manufacture 2025-06-02
UDI# (01)00385480546522(11)250602(17)261202(10)25153017

Ctl Level 1

	Spec	Pass/Fail
HDL mg/dL	31 - 53	Pass
TCHOL mg/dL	146 - 216	Pass
TGL mg/dL	48 - 124	Pass

Ctl Level 2

	Spec	Pass/Fail
HDL mg/dL	68 - 102	Pass
TCHOL mg/dL	294 - 410	Pass
TGL mg/dL	290 - 430	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

6/10/2025
Date