



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA		
Description:	TEST STRIPS, LACTATE PLUS, PACKAGED		
Catalog Number:	40813		
Nova Biomedical P/N	40813		
<u>Target Range</u>	<u>CV (%)</u>	<u>SD (mmol/L)</u>	
0.8-2.5 mmol/L	≤ 6.5%	±0.3	PASS / FAIL
3.0-5.0 mmol/L	≤ 5.5%	±0.5	PASS / FAIL
10-14 mmol/L	≤ 5.0%	±1.2	PASS / FAIL
Qty:	15,120		
Lot No.	3725153249		
Mfg. Date:	2025-06-02		
Expiration Date:	2027-06-02		
Unique Device Identifier (UDI):	(01)20385480 <u>408138</u> (11) <u>250602</u> (17) <u>270602</u> (10) <u>3725153249</u>		

Mark Broome
 Quality Control Manager

6/9/25
 Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454