



**CERTIFICATE OF ANALYSIS**

Nova StatStrip Glucose Linearity Kit  
Lot Number 0525135990

Ref Number 42173  
Exp Date 2027-11-15  
Date of Manufacture 2025-05-15  
UDI# (01)00385480421737(11)250515(17)271115(10)0525135990

Level 1 L/N 0525135301  
Level 1 Exp Date 2027-11-15  
Spec      Mean Result  
Glucose mg/dL 16 - 22      20

Level 2 L/N 0525135302  
Level 2 Exp Date 2027-11-15  
Spec      Mean Result  
Glucose mg/dL 57 - 65      62

Level 3 L/N 0525135303  
Level 3 Exp Date 2027-11-15  
Spec      Mean Result  
Glucose mg/dL 105 - 125      115

Level 4 L/N 0525135304  
Level 4 Exp Date 2027-11-15  
Spec      Mean Result  
Glucose mg/dL 275 - 325      302

Level 5 L/N 0525135305  
Level 5 Exp Date 2027-11-15  
Spec      Mean Result  
Glucose mg/dL 470 - 540      503

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. Analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

6/18/2025  
Date