



**CERTIFICATE OF ANALYSIS**

StatSensor Creatinine Control Solution Level 1  
Lot Number 5025149241

Ref Number 43921  
Exp Date 2027-05-29  
Date of Manufacture 2025-05-29  
UDI# (01)00385480439213(11)250529(17)270529(10)5025149241

	<u>Spec</u>	<u>Mean Result</u>
Creatinine mg/dL	0.90 - 1.10	0.99

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

6/16/2025  
Date