



**CERTIFICATE OF ANALYSIS**

STP pHox Ultra/CCX Lactate Membrane  
Lot Number 25150006

Part Number 48554  
Exp Date 2026-08-30  
Date of Manufacture 2025-05-30  
UDI# (01)00385480485548(11)250530(17)260830(10)25150006

	<u>Spec*</u>	<u>Mean Result</u>
Slope	≥ 15.00	37.33
pHox Ultra/CCX Ctl L4	1.7 - 2.3	1.9
pHox Ultra/CCX Ctl L5	6.3 - 7.7	6.9
Lin Soln Set A Level 1	14.0 - 20.0	18.6

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

6/17/2025  
Date