



CERTIFICATE OF ANALYSIS

STP pHox Reference Sensor
Lot Number 25160025

Ref Number 21520
Exp Date 2026-12-09
Date of Manufacture 2025-06-09
UDI# (01)00385480215206(11)250609(17)261209(10)25160025

	<u>Spec</u>	<u>Pass/Fail</u>
Na Slope	9.00 - 11.00	Pass
ADT 4	>= 220	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector

Title

6/19/2025

Date