



CERTIFICATE OF ANALYSIS

STP Prime+ Cal-Cartridge 100 Sample w. Creat/BUN
Lot Number 25161042

Ref Number 57831
Exp Date 2026-09-11
Date of Manufacture 2025-03-11

UDI# (01)00385480578318(11)250311(17)260911(10)25161042

Table with 3 columns: Cal A L/N 25071034, Cal B L/N 25071035, Cal C L/N 25071037. Rows include HCO3 mmol/L, Na mmol/L, K mmol/L, Cl mmol/L, Ca mmol/L, Mg mmol/L, Glu mg/dL, Lac mmol/L, BUN mg/dL, Creat mg/dL, and tHb g/dL with Spec, Pass/Fail, and Assigned Value\* Result columns.

Table with 3 columns: Cal D L/N 25071036, Flush L/N 25141061, Reference L/N 25070037. Rows include pH @ 37C, PO2, and KCl with Spec, Pass/Fail, and Assigned Value\* Result columns.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Handwritten signature

Approval

QC Inspector
Title

6/18/2025
Date