



CERTIFICATE OF ANALYSIS

STAT PROFILE pHox Plus Glucose Sensor
Lot Number 25163019

Part Number 27809
Exp Date 2027-06-12
Date of Manufacture 2025-06-12

UDI# (01)00385480278096(11)250612(17)270612(10)25163019

	<u>Spec*</u>	<u>Pass/Fail</u>
Slope	11.00 - 45.00	Pass
pHOx Ctl L3	75-89	Pass
Set A L1	≥ 470	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi M

Approval

QC Inspector
Title

6/24/2025
Date