



CERTIFICATE OF ANALYSIS

Prime ES Comp plus Cal Cartridge 400 Sample
Lot Number 25160048

Ref Number 62228
Exp Date 2026-11-16
Date of Manufacture 2025-05-16
UDI# (01)00385480622288(11)250516(17)261116(10)25160048

| Cal A L/N 25139020 | | | Cal B L/N 25139022 | | | Reference L/N 25136036 | | |
|--------------------|---------------|------------------|--------------------|--------------|------------------|------------------------|-------------|------------------|
| | <u>Spec</u> | <u>Pass/Fail</u> | | <u>Spec</u> | <u>Pass/Fail</u> | | <u>Spec</u> | <u>Pass/Fail</u> |
| Na mmol/L | 134.1 - 135.9 | Pass | Na mmol/L | 76.5 - 77.5 | Pass | KCl mol/L | 1.8 - 2.2 | Pass |
| K mmol/L | 4.02 - 4.08 | Pass | K mmol/L | 9.80 - 10.20 | Pass | | | |
| Cl mmol/L | 114.1 - 115.9 | Pass | Cl mmol/L | 44.6 - 45.5 | Pass | | | |
| iCa mmol/L | 0.99 - 1.01 | Pass | iCa mmol/L | 1.98 - 2.02 | Pass | | | |
| iMg mmol/L | 0.39 - 0.49 | Pass | iMg mmol/L | 1.48 - 1.52 | Pass | | | |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

06/25/2025
Date