



CERTIFICATE OF ANALYSIS

STP Prime VET ES Cal Cartridge
Lot Number 25160044

Ref Number 56329
Exp Date 2026-11-16
Date of Manufacture 2025-05-16

Cal A	L/N 25139020		Cal B	L/N 25139022		Reference L/N	25136036
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>	<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	134.1 - 135.9	Pass	Na mmol/L	76.5 - 77.5	Pass	KCl	1.8 - 2.2
K mmol/L	4.02 - 4.08	Pass	K mmol/L	9.8 - 10.2	Pass		Pass
Cl mmol/L	114.1 - 115.9	Pass	Cl mmol/L	44.6 - 45.5	Pass		
iCa mmol/L	0.99 - 1.01	Pass	iCa mmol/L	1.98 - 2.02	Pass		
iMg mmol/L	0.39 - 0.49	Pass	iMg mmol/L	1.48 - 1.52	Pass		

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

6/25/2025

Date