



CERTIFICATE OF ANALYSIS

Nova Lactate Plus Control Solution Level 2
Lot Number 3825121302

Ref Number 40814
Exp Date 2027-11-01
Date of Manufacture 2025-05-01
UDI# (01)00385480408141(11)250501(17)271101(10)3825121302

	<u>Spec</u>	<u>Mean Result</u>
Lactate mmol/L	4.4 - 5.0	4.5

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi M

Approval

QC Inspector

Title

6/25/2025

Date