



CERTIFICATE OF ANALYSIS

STP pHox Plus Cal Cartridge
Lot Number 25164015

Ref Number 27613
Exp Date 2026-11-22
Date of Manufacture 2025-05-22
UDI# (01)00385480276139(11)250522(17)261122(10)25164015

Table with columns for Cal A, B, C, D and Reference, listing analytes (Na, HCO3, K, Cl, Ca, Glu, pH, KCl) and their respective specifications and pass/fail status.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

6/24/2025
Date