



CERTIFICATE OF ANALYSIS

STP pHox Plus Calibrator Cartridge C
Lot Number 25164014

Ref Number 27612
Exp Date 2026-11-22
Date of Manufacture 2025-05-22
UDI# (01)00385480276122(11)250522(17)261122(10)25164014

Cal A L/N 25142055

Cal B L/N 25142056

Cal C L/N 25142057

	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	142 - 144	Pass	HCO3 mmol/L	39.2 - 40.8	Pass	pH @ 37°C	7.326 - 7.332	Pass
HCO3 mmol/L	19.6 - 20.4	Pass				Na mmol/L	72.5 - 73.5	Pass
K mmol/L	4.03 - 4.09	Pass				K mmol/L	10.04 - 10.20	Pass
Cl mmol/L	97.2 - 98.8	Pass				Cl mmol/L	52.6 - 53.4	Pass
Ca mmol/L	1.01 - 1.03	Pass				Glu mg/dL	198 - 202	Pass
Glu mg/dL	39.6 - 40.4	Pass						

Cal D L/N 25142058

Reference L/N 25142054

	<u>Spec</u>	<u>Pass/Fail</u>
pH @ 37C	6.782 - 6.788	Pass
Ca mmol/L	2.01 - 2.07	Pass
Glu mg/dL	990 - 1010	Pass

	<u>Spec</u>	<u>Pass/Fail</u>
KCl	1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

06/24/2025
Date