



CERTIFICATE OF ANALYSIS

STP Prime Plus Calibrator Cartridge 500 Sample
Lot Number 25174050

Ref Number 57829
Exp Date 2026-12-05

Date of Manufacture 2025-06-05

UDI# (01)00385480578295(11)250605(17)261205(10)25174050

Table with 3 columns: Cal A L/N 25162011, Cal B L/N 25163037, Cal C L/N 25162013. Rows include HCO3, Na, K, Cl, Ca, Mg, Glu, Lac, BUN and tHb g/dL with Spec, Pass/Fail, Assigned Value*, and Result columns.

Table with 3 columns: Cal D L/N 25162014, Flush L/N 25156037, Reference L/N 25162015. Rows include pH @ 37C, PO2, KCl with Spec, Pass/Fail, Assigned Value*, and Result columns.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval signature: Awishi M

QC Inspector
Title

7/2/2025
Date