



CERTIFICATE OF ANALYSIS

STP Prime Plus Calibrator Cartridge 500 Sample
Lot Number 25156032

Ref Number 57829
Exp Date 2026-11-20

Date of Manufacture 2025-05-20

UDI# (01)00385480578295(11)250520(17)261120(10)25156032

Table with 3 columns: Cal A L/N 25140079, Cal B L/N 25140080, Cal C L/N 25140081. Rows include HCO3 mmol/L, Na mmol/L, K mmol/L, Cl mmol/L, Ca mmol/L, Mg mmol/L, Glu mg/dL, Lac mmol/L, BUN mg/dL, and tHb g/dL with Spec, Pass/Fail, Assigned Value\*, and Result columns.

Table with 3 columns: Cal D L/N 25140082, Flush L/N 25141061, Reference L/N 25140098. Rows include pH @ 37C, PO2, and KCl with Spec, Pass/Fail, Assigned Value\*, and Result columns.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval signature: Awishi, M

QC Inspector
Title

7/2/2025
Date