



CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Glucose Membrane
Lot Number 25169007

Part Number 48552
Exp Date 2026-09-18
Date of Manufacture 2025-06-18
UDI# (01)00385480485524(11)250618(17)260918(10)25169007

	<u>Spec*</u>	<u>Mean Result</u>
Slope	11.00 - 45.00	34.26
pHOx Ultra/CCX Ctl L4	73 - 87	79
pHOx Ultra/CCX Ctl L5	255 - 305	287

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

6/27/2025

Date