



**CERTIFICATE OF ANALYSIS**

STP pHox Plus Sodium Sensor  
Lot Number 25168083

Ref Number 27810  
Exp Date 2026-12-17  
Date of Manufacture 2025-06-17  
UDI# (01)00385480278102(11)250617(17)261217(10)25168083

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.00 - 10.50	Pass
Std C mV	-80.00 - +80.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

7/2/2025

Date