



CERTIFICATE OF ANALYSIS

STP® pHox Ultra™ / CCX™ Creatinine Sensor
Lot Number 25167042

Part Number 48595
Exp Date 2027-06-16
Date of Manufacture 2025-06-16
UDI# (01)00385480485951(11)250616(17)270616(10)25167042

	<u>Specs</u>	<u>Pass/Fail</u>
Slope	≥ 5.00	Pass
pHOx Ctl L4	0.7 - 1.3	Pass
pHOx Ctl L5	5.2 - 7.2	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

07/02/2025

Date