



CERTIFICATE OF ANALYSIS

BioProfile FLEX 2 pH/Gas Reference Sensor
Lot Number 25177036

Part Number 57467
Exp Date 2026-12-26
Date of Manufacture 2025-06-26

| | <u>Spec*</u> | <u>Pass/Fail</u> |
|----------------|---------------|------------------|
| pH Slope | 8.70 - 11.80 | Pass |
| BioPro Ctl L 1 | 6.630 - 6.730 | Pass |
| BioPro Ctl L 3 | 7.322 - 7.422 | Pass |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to test specifications.

* Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector

Title

07/02/2025

Date