



CERTIFICATE OF ANALYSIS

Stat Profile® pHox Ultra™ / CCX™ SO2 Sensor
Lot Number 25168077

Part Number 48591
Exp Date 2028-06-17
Date of manufacture 2025-06-17
UDI#(01)00385480485913(11)250617(17)280617(10)25168077

	<u>Spec</u>	<u>Mean Result</u>	<u>Pass/Fail</u>
Slope	7.00 - 14.00	9.23	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Z. V. Aramwala
Approval

QC Inspector
Title

7/3/2025
Date