



**CERTIFICATE OF ANALYSIS**

Nova StatStrip Glucose Control Solution Level 1

Nova Ref Number 41741

Lot Number 0425150301

Exp Date 2027-11-30

Date of Manufacture 2025-05-30

UDI # (01)00385480417419(11)250530(17)271130(10)0425150301

|               | Spec    | Mean Result |
|---------------|---------|-------------|
| Glucose mg/dL | 57 - 65 | 62          |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

7/9/2025

Date