



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 200 Sample
Lot Number 25168065

Ref Number 53360
Exp Date 2026-12-06
Date of Manufacture 2025-06-06
UDI# (01)00385480533607(11)250606(17)261206(10)25168065

Cal A	L/N	25162044	
		<u>Assigned Value*</u>	<u>Result</u>
pH		7.349	N/A
pCO2 mmHg		25.3	N/A

Cal B	L/N	25160066	
		<u>Spec</u>	<u>Pass/Fail</u>
pH		6.837 - 6.843	Pass
pO2 mmHg		167.0	N/A

Flush	L/N	25157019	
		<u>Assigned Value*</u>	<u>Result</u>
pCO2 mmHg		52.4	N/A
pO2 mmHg		101.0	N/A

Reference	L/N	25160058	
		<u>Spec</u>	<u>Pass/Fail</u>
KCl		1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

06/30/2025
Date