



CERTIFICATE OF ANALYSIS

STP Prime+ Cal-Cartridge 400 Sample w. Creat/BUN
Lot Number 25147054

Ref Number 57834
Exp Date 2026-11-09

Date of Manufacture 2025-05-09

UDI# (01)00385480578349(11)250509(17)261109(10)25147054

Table with 3 columns: Cal A L/N 25129026, Cal B L/N 25129027, Cal C L/N 25129028. Rows include HCO3 mmol/L, Na mmol/L, K mmol/L, Cl mmol/L, Ca mmol/L, Mg mmol/L, Glu mg/dL, Lac mmol/L, BUN mg/dL, Creat mg/dL, pH, tHb g/dL, and Assigned Value\*/Result.

Table with 3 columns: Cal D L/N 25129029, Flush L/N 25132016, Reference L/N 25136014. Rows include pH @ 37C, PO2, and KCl.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

7/14/2025
Date