



CERTIFICATE OF ANALYSIS

BioPro Flex2 ESM Reagent Cartridge
Lot Number 25168073

Ref Number 56227
Exp Date 2026-06-02

Solution S L/N 25153036

| | <u>Spec</u> | <u>Pass/Fail</u> |
|----|---------------|------------------|
| pH | 6.900 - 7.100 | Pass |

Solution DP L/N 25153041

| | <u>Spec</u> | <u>Pass/Fail</u> |
|---------|-------------|------------------|
| NaOCl % | 0.43 - 0.73 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

7/10/2025

Date